

URINE CSF PROTEIN (2-8°C)

CATALOGUE NUMBER	KIT SIZE (ML)
MPRUCP1	2x50ml / 1x5ml

Intended Use:

For In Vitro diagnostic use by trained professionals only.

This reagent is intended for the quantitative determination of total protein in urine and cerebrospinal fluid.

Clinical Significance:

Urine does not contain any protein in healthy individuals. In the kidney the glomeruli prevent the passage of protein from the blood to the glomerular filtrate. Glomerular injury causes increased permeability to plasma proteins, resulting in proteinuria, which refers to the presence of protein in the urine. A persistent finding of proteinuria is the single most important indication of renal disease. Elevated concentrations of protein in CSF can be caused by infections and intracranial pressure.

Test Principle:

Proteins react in acid solution with pyrogallol red and molybdate to form a coloured complex. The intensity of the colour formed is proportional to the protein concentration in the sample and can be measured spectrophotometrically.

Reagent Composition

REAGENT	COMPONENT	CONCENTRATION
Reagent R1	Succinate Buffer	50 mmol/l
	Sodium Molybdate	0.04 mmol/l
	Pyrogallol Red	0.05 mmol/l
Urine/CSF Protein Standard	Albumin/Globulin aqueous primary standard	1000 mg/l

Reagent Preparation and Stability:

R1: Liquid, ready to use

Standard: Liquid, ready to use

R1 and Standard are stable to the expiry date when stored unopened at $2-8^{\circ}$ C. Once opened store tightly closed at $2-8^{\circ}$ C and protect from light and contamination. Do not use reagents after the expiry date. Dispose of reagents carefully in line with local guidelines.

Sample Collection, Preparation and Stability:

Single urine specimen or 24 hour collection samples may be used, but collected in containers without preservative.

CSF collected as standard. Any haemolysis present invalidates the sample for protein analysis.

Urine and CSF samples are stable for 48 hours at 2 - 8°C.

Assay Procedure

Assay Procedure:	
WAVELENGTH 578 – 600 nm	
TEMPERATURE	25-37°C
CUVETTE	1cm Path Length
BLANK	Reagent Blank

	Blank	Standard	Sample	
Sample	-	-	20 μΙ	
Standard	-	20 μl -		
Reagent	1000 μΙ	1000 μl 1000 μl		

Mix and incubate for 10 minutes at assay temperature. Read the absorbance (Δ Abs) of Sample/Standard against the Reagent Blank. The colour is stable for 30 minutes.

Calculation:

Urine 24h:

Concentration (mg Protein/24h) = <u>\textit{\Delta}\text{Abs Sample}\text{ x Concentration of Standard x vol (L) urine \text{\Delta}\text{\Delta}\text{ bs Standard}</u>

CSF:

Concentration (mg/L Protein) = Abs Sample x Concentration of Standard

Performance Characteristics:

Measuring range:

9.4 - 4000 mg/l.

Where results are above the linearity limit, dilute samples in a ratio of 1+3 with Normal saline, re-assay and multiply the result by 4.

Interferences:

Criterion: Recovery within +/- 10% of initial value.

Icterus: No significant interference up to 20 mg/dl of Bilirubin.

Haemolysis interferes with the assay.

Limitations:

Protein U&CSF Standard: handle carefully as this is prone to contamination. Use clean disposable tips for dispensing the calibrator supplied.

The result from this test should not be used as the sole criteria for the diagnosis of renal disease or disorders associated with elevated protein in CSF, a confirmed

diagnosis should only be made by a physician after all clinical and laboratory findings have been evaluated.

Reference Range:

Urine (24h)	< 150 mg/l
CSF	Adults 150 – 450 mg/l

Each laboratory should establish its own mean reference range according to the

Automated systems:

Contact Prestige Diagnostics Technical Department for applications on a wide range of automated analysers.

References:

- Orsonneau JL et al. An improved Pyrogallol Red-Molybdate method for determining total urinary protein. Clin Chem 1989; 35: 2233-2236.
- Koller A. Total serum protein. Kaplan A. et al. Clin Chem The C.V Mosby Co. St Louis. Toronto. Princeton 1984; 1316-1324 and 418.
- Bablok W. et al. A General Regression Procedure for Method Transformation. Clin Chem Clin Biochem 1988: 26: 783-790.
- Tietz NS (Ed) Clinical Guide to Laboratory tests, 3rd ed 1995. PA: W B Saunders; 1995: 518-522.

REF	Catalogue number		Temperature limitation
[]i	Consult instructions for use	LOT	Batch code
IVD	In vitro diagnostic medical device	X	Use by Date
***	Manufacturer	X.	Keep away from sunlight

 $C \in$

Prestige Diagnostics UK Ltd 40 Ballymena Business Centre, Galgorm, Co. Antrim, BT42 1FL, United Kingdom. Tel: +44 (0) 28 2564 2100